


SOLUTION CHANGE

DataMaster DMT: 100162
Location: SOUTH BURLINGTON PD
Date: 09/02/2016
Performed by: SARAH SUPERNEAU



Accuracy and Precision Check

Concentration = 0.098 g/210L
Lot # = 16-04-100
Range = 0.093 - 0.103
Average = 0.097 g/210L
Std Dev = 0.0013

RCD 9/2/16

Simulator Temperature: 33.8°C

Performed by

Sarah Supneau

Date

09/02/2016